

# CERTIFICATE

Choice Calibration Certificate#:12.2434

Firm:FASTEC SERVICES  
Addr:469 LINDBURG STREET  
City:CHARLESTON  
State:SC, 29412

Cal Date:03/06/2012

Choice #: F06-015

Mfg:SALEM BALL

Mod:4.50 MM

STEEL BALLS, SS316

A/N:

S/N:LOT 47165

Dept:

Loc :

Condition

Received:In Tolerance

Returned:In Tolerance

Procedure Used

QTY20+/- .005MM

Accuracy:GRADE:100

Environment

Temp.:20°C

Humid:40%

Reference Standards

A/N | Mfg. | Model | Cal.Due Date

Accuracy

Uncertny

Traceability

Number

Cal

Ratio

103 MTI 516-901 01/04/4013 GRADE 1 LD7-0204/172-10 >=4:1

110 SOKKI 225-002 02/01/2013 0.00500% 681/279723-10 >=4:1

Calibration Data: As Found & As Left.

VERIFIED QTY: 20, 4.50 MM BALLS WITHIN +/- .005 MM.

Calibration History:

No Histories on File

Choice Calibration Certifies this Instrument meets the stated accuracy with test results relating to this instrument using Manufacturer / Industry Techniques, or Customer Requirements. Services designed to comply with as required: ISO 9001:2008, ISO 10012 ISO 17025:2005 ANSI/NCSL Z540 former MIL-STD-45662A Using Certified Standards Traceable to N.I.S.T. as Required Uncertainties Estimated at a 95% Confidence Level (k=2).

No reproduction except in full without written approval.

Choice Calibration Company 17411 Irvine Blvd. Unit C  
Tustin, CA 92780 Ph: 714-730-1670, Fax: 714-730-7818

Q.C.  
11

QA Approval, By:Ken Meadows

Issued Date:03/06/2012